



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/596,034

01/31/2007

David Bassin

3869/029 US

1908

22440

7590

11/09/2010

GOTTLIEB RACKMAN & REISMAN PC

270 MADISON AVENUE

8TH FLOOR

NEW YORK, NY 10016-0601

EXAMINER

LOUIS, LATOYA M

ART UNIT

PAPER NUMBER

3771

MAIL DATE

DELIVERY MODE

11/09/2010

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/596,034	Applicant(s) BASSIN, DAVID	
	Examiner LaToya Louis	Art Unit 3771	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 16 August 2010.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 112-134 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 112-134 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

Art Unit: 3771

DETAILED ACTION

1. This office action is responsive to the amendment filed 8/16/2010. As directed by the amendment, claims 112-134 have been amended, claims 1-111 have been previously cancelled, and no new claims have been added. Thus claims 112-134 are currently pending.

Continued Examination Under 37 CFR 1.114

2. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 8/16/2010 has been entered.

Claim Rejections - 35 USC § 112

3. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

4. Claims 113-120 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Regarding claim 113, the limitation “the amplitude” lacks antecedent basis.

Regarding claim 117, the limitations “the current breathing cycle” and “the amplitude” lack antecedent basis.

Claims 114-116 and 118-120 are rejected for their dependency on a rejected claim.

Claim Rejections - 35 USC § 102

5. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

6. Claims 112-115, 121, 122, 125, and 128-134 are rejected under 35 U.S.C. 102(b) as being anticipated by Berthon-Jones (EP 1,129,742 A2).

Regarding claim 112, Berthon-Jones teaches in fig. 2 an apparatus for providing ventilatory pressure support to a patient comprising a control mechanism (i.e. 16) for deriving two calculated errors (page 6 lines 45-56 discloses a formula, $5 \text{ lfl-V}_{\text{TGT}}$, for calculating at least two errors values. As disclosed the first calculated error results when $5 \text{ lfl-V}_{\text{TGT}} = \text{negative}$ indicating that the patient is under ventilated and requires more and sudden ventilatory support. A second calculated error results when $5 \text{ lfl-V}_{\text{TGT}} = \text{positive}$ indicating the subject is over ventilated and requires less and smoother ventilatory support. In addition, multiple calculated error values will result from $5 \text{ lfl-V}_{\text{TGT}}$ depending on the measured values of airflow. [0031] and [0042] disclose large negative error values as alternative first calculated error indicating that the patient is grossly under ventilated and requires faster ventilator support and small negative error values as alternative second calculated error indicating that the patient is mildly under ventilated and requires smoother ventilator support) each of which is a function of the same target ventilation value (V_{TGT}) and a respective one of two patient ventilation measures the two patient ventilation measures having respective relatively fast and relatively slow speeds of response to

Art Unit: 3771

the calculated errors, (i.e. a measured airflow value that is grossly under V_{TGT} as first ventilation measure having $K=0$ and a measured airflow value that is over or only slightly below V_{TGT} as second ventilation measure having $K=1$), the control mechanism further deriving two control responses of pressure to respective ones of the two calculated errors (As indicated in [0029] and shown in figs. 3 and 4, The first control response is when $\Pi(\Phi)$ becomes a square wave causing a sudden and efficient pressure response when $K=0$ as a result of the a first calculated error and the second control response is when $\Pi(\Phi)$ becomes a smoother wave causing a smoother less efficient pressure response when $K=1$ as a result of the second calculated error) and combining the two control responses to produce an overall control response that increasingly favors the first control response (square wave) as the first ventilation measure (airflow below or grossly below V_{TGT}) becomes increasingly less than the target ventilation value (page 7 lines 1-25 and [0038]), and a ventilator (i.e. 19) responsive to the overall control response for controlling the level of pressure of air delivered to said patient (page 4 lines 10-14).

Regarding claim 113, Berthon-Jones discloses that each of the two control responses is a function of the amplitude and sign of the respective one of the calculated errors (page 6 lines 25-26 and 45-55) so that the first control response (square wave) is more vigorous than the second control response (smooth wave) ([0022]).

Regarding claims 114, 121, 129, and 130 Berthon-Jones discloses that the degree of control and gain exercised by the ventilator servo control mechanism increases with the magnitudes of the two calculated errors (page 7 lines 10 and 25-35).

Art Unit: 3771

Regarding claims 115, 122, and 131 Berthon-Jones discloses that the degree of control/servo exercised by the ventilator is greater (has a greater gain G) for calculated errors below the target value (page 7 lines 25-31 and the code found on lines 40-51 indicate that for calculated errors above V_{TGT} only amplitude A decreases but for calculated errors below V_{TGT} K is incremented towards 1, making the wave more square, and the amplitude A increases thereby providing a greater degree of control).

Regarding claim 125, Berthon-Jones discloses that the control mechanism further determines the phase of the current breathing cycle ([0020]) and adjusts the overall control response to be a function of the amplitude at the determined phase of the current breathing cycle of an amplitude-versus-phase template that is appropriate for a normal breathing cycle ([0019] and page 5 lines 21-36 disclose that the control response, being dependent on K , is a function of amplitude A at a particular breathing phase of a pressure waveform template $\Pi(\Phi)$.

Regarding claim 128, Berthon-Jones discloses that each of the calculated errors is a clipped integral of the respective patient ventilation measure minus said target value ([0032]).

Regarding claim 132, Berthon-Jones discloses that the gain is varied more aggressively for conditions of hypoventilation than for conditions of hyperventilation (page 6 lines 46-55).

Regarding claim 133, Berthon-Jones discloses that the ventilator is flow-triggered and phase cycled (page 4 lines 46-54).

Regarding claim 134, Berthon-Jones discloses that the ventilator withdraws ventilation support more gradually when the patient is over-ventilated than when the patient is under-ventilated (page 7 lines 25-31 and the code found on lines 40-51 indicate that for calculated

Art Unit: 3771

errors above V_{TGT} only amplitude A decreases while K remains at 0 but for calculated errors below V_{TGT} K is incremented towards 1, making the wave more square and abrupt while also increasing the amplitude).

Claim Rejections - 35 USC § 103

7. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

8. Claims 116, 117, 123, and 124 are rejected under 35 U.S.C. 103(a) as being unpatentable over Berthon-Jones in view of the article entitled "An Adaptive Lung Ventilation Controller" of record by Laubscher et al.

Regarding claims 116, 123, and 124, Berthon-Jones teaches target value (V_{TGT}) but doesn't teach that the target value is an alveolar ventilation that takes into account the patient's anatomical or physiologic dead space. However, Laubscher teaches an apparatus (fig. 2) for providing ventilatory assistance wherein the target value " V'_{ga} " is an alveolar ventilation (page 51, col. 2, 4th paragraph) that takes into account the patient's anatomical or physiologic dead space " VD " (page 52 col. 1 paragraph 2 and in formula (2) on page 51 teaches that the target value " V'_{ga} " is a desired ventilation entered by a user as in that takes into account the patient's physiologic dead space VD). It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the target value of Berthon-Jones to take into account the

Art Unit: 3771

physiologic dead space as taught by Laubscher to provide more accurate ventilation in people with lung disease

Regarding claim 117, Berthon-Jones discloses that the control mechanism further determines the phase of the current breathing cycle ([0020]) and adjusts the overall control response to be a function of the amplitude at the determined phase of the current breathing cycle of an amplitude-versus-phase template that is appropriate for a normal breathing cycle ([0019] and page 5 lines 21-36 disclose that the control response, being dependent on K, is a function of amplitude A at a particular breathing phase of a pressure waveform template $\Pi(\Phi)$.

9. Claims 118-120 are rejected under 35 U.S.C. 103(a) as being unpatentable over Berthon-Jones in view of Laubscher et al., as applied to claim 117 above, and further in view of Berthon-Jones (WO 01/19440 A1) of record hereinafter Berthon-Jones '40.

Regarding claims 118 and 119, Berthon-Jones discloses that the control mechanism determines the phase of the current breathing cycle by applying a set of fuzzy logic rules ([0020]) but does not specifically disclose that the phase is determined by relating respiratory airflow and its rate of change to different phases of a normal breathing cycle. However, Berthon-Jones '40 teaches that the phase is determined by relating respiratory airflow and its rate of change to different phases of a normal breathing cycle (Page 4 lines 4-8). It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the control mechanism of Berthon-Jones to determine the phase using respiratory airflow

Art Unit: 3771

and its derivative as taught by Berthon-Jones '40 to provide an accurate and less complicated/easier to obtain means of determining the phase.

Regarding claim 119, Berthon-Jones discloses that the overall control response is a clipped integral of a function of both of the calculated errors (page 6 line 30).

10. Claims 126 and 127 are rejected under 35 U.S.C. 103(a) as being unpatentable over Berthon-Jones in view of Berthon-Jones '40.

Regarding claims 126 and 127, Berthon-Jones discloses that the control mechanism determines the phase of the current breathing cycle by applying a set of fuzzy logic rules ([0020]) but does not specifically disclose that the phase is determined by relating respiratory airflow and its rate of change to different phases of a normal breathing cycle. However, Berthon-Jones '40 teaches that the phase is determined by relating respiratory airflow and its rate of change to different phases of a normal breathing cycle (Page 4 lines 4-8). It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the control mechanism of Berthon-Jones to determine the phase using respiratory airflow and its derivative as taught by Berthon-Jones '40 to provide an accurate and less complicated/easier to obtain means of determining the phase.

Response to Arguments

11. Applicant's arguments with respect to claims 112-134 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

12. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Hill (2002/0088465) discloses a method wherein the degree of IPAP is changed based on degrees of error.

13. Any inquiry concerning this communication or earlier communications from the examiner should be directed to LaToya Louis whose telephone number is (571) 270-5337. The examiner can normally be reached on Monday-Friday, 8:30am-6pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Justine Yu can be reached on 571-272-4835. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/LaToya Louis/
Examiner, Art Unit 3771
11/3/2010

/Justine R Yu/
Supervisory Patent Examiner, Art Unit 3771

Application/Control Number: 10/596,034

Page 10

Art Unit: 3771